

Ohm METRICS Resistance Meter Model SRM310 - Pocket-Sized Meter

Quickly and accurately measure the resistance of materials with the push of a button.

This SRM 310 uses parallel electrodes and concentric rings to accurately measure RTT, RTG, or resistivity for periodic verification, factory audits or test lab evaluation of product.

This meter is designed to verify incoming materials such as floor mats, table mats, bins and shelving in static controlled areas. Work surfaces, bags, wrist straps, footwear, packaging materials, garments, and ESD chairs can also be tested with the device.

The SRM310 meets periodic test requirements per Compliance Verification ESD TR53 and conforms to ANSI/EOS/ESD (S4.1, S7.1, S12.1, S2.1).



Features

- Quickly and Accurately Measure
 Surface Resistance for ESD Compliance
- Test range: 10^3-10^12ohms
- Test voltage: 10V/100V(automatic)
- Easy To Use: One Button Operation
- Accuracy: +/- 5 %

Applications:

Surface Resistance Meter is a portable battery powered checker fitted with built-in parallel electrodes that allow for tests of material surface resistance. This meter is designed for quick checks of surface resistance for ESD control applications in electronics manufacturing or handling environment.

_			
Sna	aciti	c atı	onc:
200		Lau	ons:

Test Range 10^3-10^12 **Dimensions** 5.15" H x 2.50" W x 1.8" D Test Voltage 10V/100V(Automatic Ranging) Weight 8.46 oz 9V-Battery Indicators Green: <10^5 (Conductive) Power +/-5% Yellow: 10^6-1011 (Dissipative) Accuracy Red: 10^12> (Insulating) **Probes Built-in Parallel Probes** Periodic Verification CAL1000 Meter Adjustment No adjustments required

Part Numbers:

SRM310 CAL1000 Surface Resistance Meter Periodic Verification Unit

This document is prepared for our customers as a service, and is to the best of our knowledge true and accurate. However, it is understood and agreed by the users of this document that we will accept no liability for the conclusions reached. Users of this document may therefore wish to perform additional testing before determining that products mentioned are suitable.